



Certificate of Analysis

Product Code Number 1509A-25I

Product Name COVID-19 Antigen Rapid Test Kit

Product Lot Number COV6696B

Expiration Date 2022-11

Storage Temperature 2° to 30°C

| CODE # | COMPONENT | LOT # | EXP. DATE |
|--------|-------------------|----------|-----------|
| 1509A | Cassettes | COV21849 | 2023-01 |
| 4279 | Extraction Buffer | 21850 | 2023-01 |

| | Sample Concentration | Specification | Result | Pass/Fail |
|---------------------|------------------------|---------------|----------|-----------|
| QC Functional Check | Negative Sample | Negative | Negative | Pass |
| | Low Positive Sample | Positive | Positive | Pass |
| | Medium Positive Sample | Positive | Positive | Pass |

The product listed above has been tested according to Biomerica's Quality Control Procedures and has been found to meet Biomerica's internal Quality Control Specifications.

Product Released By: N. Fernandez 14-May-21
Quality Control Date

Biomerica, Inc.
17571 Von Karman Avenue, Irvine, CA 92614, USA
Tel: 949-645-2111 Fax: 949-553-1231
www.biomerica.com, bmra@biomerica.com